
Re: Defibrillator outside Community Shop

From Maggie Burt <clerk.coddenhampc@gmail.com>

Date Tue 2025-09-23 10:55 AM

To Sarah Gregory <cllrSarahgregory@gmail.com>; JANE SOANES <jane.soanes@btinternet.com>

Cc coddenhamcllr.nickmills@gmail.com <coddenhamcllr.nickmills@gmail.com>

According to their website, AED Expert quote 2-3 working days for delivery.

Kind Regards,
Maggie

From: Sarah Gregory <cllrSarahgregory@gmail.com>

Sent: 23 September 2025 10:42

To: JANE SOANES <jane.soanes@btinternet.com>

Cc: clerk.coddenhampc@gmail.com <clerk.coddenhampc@gmail.com>; coddenhamcllr.nickmills@gmail.com <coddenhamcllr.nickmills@gmail.com>

Subject: Re: Defibrillator outside Community Shop

Jane, hi

Batteries deteriorate so I would not support buying one until we are nearer end of life for the existing battery. As the current battery is only two, I would suggest we budget to have a replacement when the battery will be 4 +.

I agree we need to replace pads on expiry. Again, do we really need spare ones? . Again, they will have an expiry date.

Maggie- please can you confirm how long delivery will take. Presumably only a day or two?

Many thanks
Kind regards
Sarah

Sent from my iPhone

On 23 Sep 2025, at 10:10, JANE SOANES <jane.soanes@btinternet.com> wrote:

Hi Maggie,

Thank you.

We will definitely need to replace the pads by October 26 as the current ones expire then.

The current battery is only 2 years old and hasn't been used. It's lifespan will reduce if

it is used.

It would be good practice to also have a spare set of pads as when they are used they need to be replaced.

It would also be good practice to have a spare battery. As when the battery fails the defibrillator would be offline and not available for use until a new battery is sourced. I would therefore propose to buy a spare battery and pads in this financial year and then get a another set of pads in the next financial year for when the current ones expire.

This obviously has an impact on our budget so **Sarah, Nick...your thoughts??**

Kind regards

Jane

Sent via BT Email App

From: Maggie Burt <clerk.coddenhampc@gmail.com>

Sent: 23 September 2025 09:31:40 BST

To: jane.soanes <jane.soanes@btinternet.com>, cllrsarahgregory@gmail.com
<cllrsarahgregory@gmail.com>, coddenhamcllr.nickmills@gmail.com
<coddenhamcllr.nickmills@gmail.com>

Subject: Re: Defibrillator outside Community Shop

Hi Jane

I've investigated the costs of replacement pads and batteries and below are the details:

Defib Shop

Pads £90+VAT

Battery £250+VAT

St Johns Ambulance

Do not supply our model of AED

AED Expert

Pads £60+VAT

Battery £225+VAT

Battery lifespan is quoted by suppliers as being between 4.5 years to 5.0 years and the pads supplied would be for both adults and children.

If you think we will need to purchase these replacements in this financial year, I will add this to the Agenda for the November meeting.

Kind Regards,
Maggie

From: jane.soanes <jane.soanes@btinternet.com>

Sent: 20 September 2025 22:59

To: clerk.coddenhampc@gmail.com <clerk.coddenhampc@gmail.com>;
cllrsarahgregory@gmail.com <cllrsarahgregory@gmail.com>;
coddenhamcllr.nickmills@gmail.com <coddenhamcllr.nickmills@gmail.com>

Subject: Defibrillator outside Community Shop

Hi Maggie,next

The defibrillator will be 3 years old next October so the PC will have to replace the pads. Also for budget purposes we need to factor in the cost of a replacement battery. Could you please investigate the following:-

1. How long is the expected lifespan of a battery if the defibrillator is not used.
2. The cost of a replacement battery.
3. The cost of replacement adult and child pads.

The defibrillator is a Beneheart C1A

Thanks

Jane